Drägerwerk AG & Co.KGaA, 23542 Lübeck

Our reference
-
Phone number
+49 451 882-6166

E-mail
hannes.molsen@draeger.com
fabian.busch@draeger.com

Vulnerability and Incident Reporting
product-security@draeger.com

Dräger Product Security Advisory

CVE-2019-11687 Security Advisory for DICOM Images

>> Dräger Medical Devices are not affected by CVE-2019-11687.

To whom it may concern,

On April 26, 2019, the security company cylera released information about a vulnerability in the DICOM file format.
None of our medical devices uses the DICOM file format, neither for export, nor for import.
Therefore Dräger Medical Devices are not affected by the aforementioned vulnerability.
For reporting incidents and potential vulnerabilities in our devices, please refer to https://static.draeger.com/security to contact the Product Security team directly.

Best regards,

Dräger Product Security

Revisions

Version	Date	Change
1	2019-06-19	Initial Document