Drägerwerk AG & Co.KGaA, 23542 Lübeck

Our reference
-
Phone number
+49 451 882-6166

Vulnerability and Incident Reporting
product-security@draeger.com

Dräger Product Security Advisory PSA-21-109-01

Open Source Network Stack Vulnerabilities "NAME:WRECK"

>> Dräger Medical Devices are not affected by "NAME:WRECK".

To whom it may concern,

On April 14, 2021, the cybersecurity company forescout disclosed a total of 9 security vulnerabilities in the DNS implementations of 4 different open source network stacks[1]. The advisory describes the affected network stacks and the individual security vulnerabilities. None of Dräger medical devices or the accompanying infrastructure utilizes the affected network stacks in vulnerable versions on its external communication interfaces.

For reporting incidents and potential vulnerabilities in our devices, please refer to https://static.draeger.com/security to contact the Product Security team directly.

Best regards,

Dräger Product Security

[1] https://www.forescout.com/research-labs/namewreck/

Revisions

Version	Date	Change
1	2021-04-19	Initial Document

Drägerwerk AG & Co. KGaA
Moislinger Allee 53-55
23558 Lübeck, Germany
Postal address:
23542 Lübeck, Germany
Tel +49 451 882-0
Fax +49 451 882-2080
info@draeger.com
www.draeger.com
VAT no. DE135082211

Bank details:
Commerzbank AG, Lübeck
IBAN: DE95 2304 0022 0014 6795 00
Swift-Code: COBA DE FF 230
Sparkasse zu Lübeck
IBAN: DE15 2305 0101 0001 0711 17
Swift-Code: NOLADE21SPL

Registered office: Lübeck
Commercial register:
Local court Lübeck HRB 7903 HL
General partner: Drägerwerk Verwaltungs AG
Registered office: Lübeck
Commercial register:
Local court Lübeck HRB 7395 HL

Chairman of the Supervisory Board
for Drägerwerk AG & Co. KGaA
and Drägerwerk Verwaltungs AG:
Stefan Lauer
Executive Board:
Stefan Dräger (chairman)
Rainer Klug
Gert-Hartwig Lescow
Dr. Reiner Piske
Anton Schrofner