Drägerwerk AG & Co.KGaA, 23542 Lübeck

Our reference
-
Phone number
+49 451 882-6166

E-mail
hannes.molsen@draeger.com
detlef.koeble@draeger.com
matt.mckenna@draeger.com

Vulnerability and Incident Reporting
product-security@draeger.com

Dräger Product Security Advisory PSA-20-171-2

Treck Network Stack Security Vulnerabilities “Ripple20”

>> Dräger Medical Devices are not affected by “Ripple20”

Legal Notice

This Product Security Advisory is based on all our findings that we had at the time of publication. However, the facts of the case are being reviewed on an ongoing basis and it is possible that this may result in changed assessments or that assessments contained in this advisory may turn out to be incorrect. We also reserve the right to change or revoke any recommendations. In addition, differences may result from the circumstances of the individual case on site. This information is naturally not available to us and has not been taken into account. Dräger can therefore accept no responsibility that the information presented here is conclusive or comprehensively correct for you. Therefore, please check carefully to what extent deviations can arise for your individual case. If necessary, you will be informed about new findings through following advisories.

To whom it may concern,

On June 16, 2020, the security company JSOF released information about several vulnerabilities in the widely used Treck Network Stack.

None of Dräger medical devices uses this network stack, therefore Dräger Medical Devices are not affected by the vulnerabilities.

For reporting incidents and potential vulnerabilities in our devices, please refer to https://static.draeger.com/security to contact the Product Security team directly.

Best regards,

Dräger Product Security

[Update: 03.07.2020]

Dräger tests revealed no problems with ICM in conjunction with a Digi Connect ES 4 Version 2.24 from Digi International GmbH.

Revisions

VERSION	DATE	CHANGE
1.0	2020-06-19	Initial Document
2.0	2020-07-03	[Update: 03.07.2020]

Drägerwerk AG & Co. KGaA
Moislinger Allee 53-55
23558 Lübeck, Germany
Postal address:
23542 Lübeck, Germany
Tel +49 451 882-0
Fax +49 451 882-2080
info@draeger.com
www.draeger.com
VAT no. DE135082211

Bank details:
Commerzbank AG, Lübeck
IBAN: DE95 2304 0022 0014 6795 00
Swift-Code: COBA DE FF 230
Sparkasse zu Lübeck
IBAN: DE15 2305 0101 0001 0711 17
Swift-Code: NOLADE21SPL

Registered office: Lübeck
Commercial register:
Local court Lübeck HRB 7903 HL
General partner: Drägerwerk Verwaltungs AG
Registered office: Lübeck
Commercial register:
Local court Lübeck HRB 7395 HL

Chairman of the Supervisory Board
for Drägerwerk AG & Co. KGaA
and Drägerwerk Verwaltungs AG:
Stefan Lauer
Executive Board:
Stefan Dräger (chairman)
Rainer Klug
Gert-Hartwig Lescow
Dr. Reiner Piske
Anton Schrofner